

PRODUCT DESCRIPTION

Epinephrine is a lyophilized preparation of adrenalin. The working concentration of the reconstituted reagent is 1×10^{-3} M.

INTENDED USE

Epinephrine is for use in routine platelet aggregation studies for the evaluation of platelet dysfunction or platelet activation.

PRINCIPLE

When added to platelet rich plasma, Epinephrine stimulates platelets to aggregate. Aggregation induced by Epinephrine is referred to as primary aggregation. Normal platelets will further respond by releasing endogenous ADP from their granules. Release of endogenous ADP results in a secondary wave of aggregation.^{8,10,11}

PRECAUTIONS

Epinephrine is for *IN-VITRO* DIAGNOSTIC USE ONLY AND NOT FOR INJECTION OR INGESTION.

MATERIALS PROVIDED

Epinephrine, 3 x 0.5mL. Store at 2° to 8° C prior to reconstitution.

MATERIALS REQUIRED BUT NOT PROVIDED

1. Platelet Aggregometer
2. Purified water (distilled, deionized or reagent grade), pH 5.3 - 7.2
3. Pipettors (0.5mL, 0.45mL, 0.05mL volumes)
4. Disposable Stir bars
5. Aggregometer cuvettes

INSTRUMENTATION

Epinephrine will perform as described when used on most optical platelet aggregometers.¹ Follow the manufacturer's instructions for operating the aggregometer in use.

SPECIMEN COLLECTION AND PREPARATION OF TEST SAMPLE

Refer to the current NCCLS Approved Guideline H21 A2 for detailed specimen collection and sample preparation instructions.⁶

1. PATIENT PREPARATION:

Patients should refrain from taking aspirin or medications containing aspirin, other medications and dietary supplements known to affect platelet function for 7 - 10 days prior to specimen collection. Patients should fast and avoid fatty foods and dairy products for 12 hours prior to specimen collection.⁶

2. SPECIMEN COLLECTION:

Blood collection should be performed with care to avoid stasis, hemolysis, contamination by tissue fluids, or exposure to glass. Keep specimens at room temperature.⁸

Each of the following can cause test results to be inaccurate; and affected specimens should be rejected: hemolysis, RBC contamination, lipemia, chylous, icterus, thrombocytopenia ($<75,000/\text{mm}^3$) clots in specimen, and hypofibrinogenemia. Reuse of disposable items may result in inaccurate test results.

Observe standard precautions throughout the specimen collection, sample preparation and analytical processes.^{2,3} Dispose of sharps and biological waste in accordance with laboratory policy.

Syringe Technique (recommended)⁸

- a. Use a butterfly needle for the venipuncture.
- b. Draw 9.0mL of blood into a plastic syringe. Avoid excess suction.
- c. Remove the needle from the syringe and immediately and gently dispense the blood into a plastic [polypropylene]4 tube containing 1.0mL of 0.11M Sodium Citrate anticoagulant. The ratio of blood to anti-coagulant must be 9 parts of blood to 1 part anti-coagulant.⁵
- d. Cover and invert 4-5 times gently to mix.
- e. Maintain at room temperature (15° to 28°C).

NOTE: When the patient's hematocrit is $< 30\%$ or $> 55\%$, the blood to anticoagulant volumes must be adjusted.⁴

Evacuated Collection Tube Technique.

1. Use a butterfly needle for the venipuncture.
2. Draw blood using (plastic) tubes containing 0.11M Sodium Citrate anticoagulant.
3. Gently invert 4-5 times to mix.

NOTE: When using plastic vacuum collection tubes, make sure the citrate anticoagulant is 0.11M by checking the label. Colored tops do not vary with differing citrate concentrations. Follow the manufacturer's instructions for specimen collection.

PREPARATION OF PLATELET RICH PLASMA (PRP) AND PLATELET POOR PLASMA (PPP)

1. Prepare platelet rich plasma by centrifuging the anti-coagulated blood at 150 X g for 10 minutes at room temperature (15° to 28°C).
2. Examine the plasma layer for red cells. If red cells are present, re-centrifuge at 150 X g for an additional 5 minutes.
3. Using a plastic transfer pipette, observe and carefully remove the platelet layer without disturbing the buffy coat or red cells, and transfer to a container labeled (PRP). Cap the container and allow it to stand at room temperature.
4. Prepare the platelet poor plasma by centrifuging the remaining blood specimen at 2500 x g for 20 minutes. Examine the platelet poor plasma for hemolysis, then transfer it to a plastic tube labeled PPP.
5. The platelet count of the PRP should be $250,000 \pm 50,000/\text{mm}^3$. The platelet count may be reduced by using PPP prepared from the sample.

NOTE: If using Arachidonic Acid as an agonist, do not adjust the platelet count.

RECONSTITUTION

NOTE: Reagents must be at room temperature (15° to 28°C) prior to reconstitution. Stored reagent must be brought to room temperature prior to use.

Reconstitute a vial of Epinephrine with 0.5mL purified water.

REAGENT STORAGE

The reconstituted Epinephrine is stable for 30 days when stored at 2°-8°C in its original tightly sealed container.

TEST PROCEDURE

The platelet rich plasma must be held at room temperature for at least 30 minutes prior to testing.

1. Prepare an aggregometer blank by pipetting 0.5mL platelet poor plasma into a cuvette.
2. Pipette 0.45mL platelet rich plasma into a second cuvette. Incubate at 37°C for 3 minutes and add a stir bar.
3. Set, if required, the 0% and 100% baselines according to the manufacturer's instructions for the aggregometer in use.
4. Add 0.05mL Epinephrine directly into the platelet rich plasma. Do not allow reagent to run down the wall of the cuvette. The final concentration of Epinephrine in the platelet rich plasma is 1×10^{-4} M.
5. Allow the aggregation pattern to generate for 5 minutes.

QUALITY CONTROL

Laboratories should follow generally accepted quality control practices when proficiency testing is not available.

To assure proper instrument operation and reagent performance, a control specimen should be evaluated each day that tests are performed. The control specimen should be prepared in the same manner as the test specimen. For qualitative platelet aggregation studies, the control should consist of fresh platelet rich plasma collected from a (specified and qualified) normal donor who has not ingested aspirin containing compounds within 10 days of testing and has a history of normal platelet function.

RESULTS

Typical Epinephrine aggregation patterns are illustrated in Figures 1 and 2. Epinephrine will induce two distinct waves of aggregation in normal platelet rich plasma.^{8,10,11}

EXPECTED VALUES

Expected ranges for each reagent at various concentrations used to induce platelet aggregation should be established by each laboratory, see Table 1.^{4,8,9,10}

Table 1

TYPICAL PLATELET AGGREGATION RESPONSES FOR NORMAL DONORS @ 250,000 PLATELETS/mm³ [total aggregation at 5 minutes]

	ADP	Arachidonic Acid	Collagen [Type I]	Epinephrine
Final Conc.	2.0×10^{-5} M	500µg/mL	0.19mg/mL	1.0×10^{-4} M
Lag Phase [sec]	<10	<=20	<60	0
Primary Slope	38-67	>20	35-67	7-34
Total Aggregation (%@5min)	63-89	65-90	61-99	54-101
Biphasic Aggregation	concentration dependent	NO	NO	YES
Other	May show Shape changes	All normal Donors may not Conform PLT CT~175k-300k	Do not Dilute	All normal Donors may not Conform




Fig. 1 Normal Aggregation




Fig. 2 Abnormal Aggregation

LEGEND: Results of Epinephrine induced platelet aggregation on normal and abnormal plasmas. The final concentration of Epinephrine in PRP is $1 \times 10^{-4}M$. Spike mark indicates the addition of reagent.

LIMITATIONS³

Spurious results will be observed when the platelet count of the platelet rich plasma is less than 75,000 platelets/uL.

Platelet rich plasma which has not been held at room temperature for at least 30 minutes prior to testing may yield abnormal results.

Several reports indicate that platelet rich plasma from 20-50% of the normal population will exhibit only a primary wave of aggregation in response to Epinephrine.

PERFORMANCE CHARACTERISTICS

Studies have shown that this product will perform as described prior to its expiration date when procedural and storage directions are followed.

Linearity:

Platelet aggregation induced by common agonists (ADP, Arachidonic Acid, Collagen and Epinephrine) is a nonlinear test system for the following parameters: Lag Phase, Primary Slope, Secondary Slope, biphasic response and disaggregation. The non-linearity is caused by many factors such as the reaction chemistry and instrumentation. Platelet aggregation measures a response rate or activity that is not a quantitative measure of the reactants or their concentration.

ACCURACY, PRECISION AND REPRODUCIBILITY

Accuracy

In platelet aggregation, accuracy is a relative parameter and is dependent on the test system.

Precision and Reproducibility

The limitations of platelet aggregation makes it difficult to provide typical precision or reproducibility ranges. However, there is an experienced based consensus for these parameters (see below). Each laboratory must establish its own limits for test acceptability.

Test to Test Reproducibility:	better than $\pm 7.5\%$
Instrument to Instrument Reproducibility:	better than $\pm 15\%$
Reagent Lot to Lot Variation:	better than $\pm 10.5\%$
Laboratory to Laboratory (same test system):	better than $\pm 12.5\%$

REFERENCES

- Born, GVR and Cross, MJ. The Aggregation of Blood Platelets. J. Physiol [London] 168:178, 1963.
- Centers for Disease Control and Prevention. Guidelines for Isolation Precautions in Hospitals. Centers for Disease Control and Prevention. 1996; Vol 17; 1:53 - 80.
- National Committee for Clinical Laboratory Standards. NCCLS: Protection of Laboratory Workers from Occupationally Acquired Infections: Approved Guideline. NCCLS document M29. Wayne, PA.
- McCabe-White, M and Jennings, LK. Platelet protocols: Research and Clinical Laboratory Procedure. Academic Press. London. 1999, p 35.
- Newhouse, P and Clark, C. The Variability of Platelet Aggregation., in Triplett, DA,ed. Platelet Function: Laboratory Evaluation and Clinical Application. ASCP. Chicago. 1978. p 69.
- National Committee for Clinical Laboratory Standards. NCCLS Collection, Transport and Processing of Blood Specimens Approved Guideline- Second Edition. NCCLS Document H 18-A2. Wayne, PA .
- Weiss HJ: Aspirin and platelets in drugs and hematologic reactions. Dimitov and Nodine (eds.). Grune and Stratton, New York, 1974.

- Triplett DA, Harms CS, Newhouse P, Clark C: Platelet Function.Laboratory Evaluation and Clinical Application. ASCP, 1978.
- Day HJ, Holmsen H: Laboratory tests of platelet function. Annal Clin Lab Sci, 2:63, 1972.
- Owen CA, Bowie EJW, Thompson JH: The diagnosis of bleeding disorder. Little, Brown and Co., 1975.
- Williams WJ, Beutler, E. Erslev AJ, Rundles RW: Hematology. McGraw-Hill, 1977.

PRODUCT AVAILABILITY

PRODUCT	NET CONTENTS	CATALOG NUMBER
Epinephrine	3 x 0.5mL	101311
ADP	3 x 0.5mL	101312
Arachidonic Acid	3 x 0.5mL	101297
BETA/Pak®		
(ADP, Collagen, Ristocetin)	1 x 0.5mL each	101580
Collagen	3 x 0.5mL	101562
Lyophilized Platelets	3 x 4 mL	101595
Lyophilized Platelets	1 x 10 mL	101258
PAR/Pak® II		
(ADP, Collagen, Epinephrine)	2 x 0.5mL each	101310
Ristocetin		
AggRecetin® 1.5mg/mL	15mg	100968
AggRecetin® 1.0-1.5mg/mL	15mg	100970
AggRecetin Bulk	100mg	101241
vW Factor Assay®	10 Determinations	101246
vW Factor Assay	20 Determinations	103025
vW Abnormal Control Plasma	3 x 0.5mL	101270
vW Normal Reference Plasma	3 x 0.5mL	101269
vW Normal Control Plasma	3 x 0.5mL	106426

THIS PRODUCT IS WARRANTED TO PERFORM AS DESCRIBED IN THE LABELING AND IN THE LITERATURE OF BIO/DATA CORPORATION AND BIO/DATA CORPORATION DISCLAIMS ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR ANY OTHER PURPOSE AND IN NO EVENT SHALL BIO/DATA CORPORATION BE LIABLE FOR ANY CONSEQUENTIAL DAMAGES ARISING OUT OF THE AFORESAID EXPRESSED WARRANTY.



155 Gibraltar Road, PO Box 347, Horsham, PA 19044-0347 U.S.A.
 (800) 257-3282 U.S.A. (215) 441-4000 Worldwide
 (215) 443-8820 Fax Worldwide
 E-mail: bdc@biodatacorp.com
 Internet: www.biodatacorp.com



EMERGO EUROPE, Molenstraat 15, 2513 BH The Hague, The Netherlands

